

試験結果 Test Data

管理番号: 1234567890
(Control No.)

型名 (Model No.) 7 3 5 2 A
製造番号 (Serial No.) 1 2 3 4 5 6 7 8 9

[Ach]

1. 直流電圧測定 (DC Voltage Measurement) (5 1/2 digits display, IT:100ms(SLOW1))

レンジ (Range)	試験値 (Tested)	規格下限 (Min)	測定値 (Data)	規格上限 (Max)	判定 (Pass/Fail)
200 mV	190.000 mV	189.972 mV	190.000 mV	190.028 mV	良 (Pass)
	-190.000 mV	-190.028 mV	-190.001 mV	-189.972 mV	良 (Pass)
2000 mV	1900.00 mV	1899.78 mV	1900.01 mV	1900.22 mV	良 (Pass)
	-1900.00 mV	-1900.22 mV	-1900.01 mV	-1899.78 mV	良 (Pass)
20 V	19.0000 V	18.9968 V	19.0000 V	19.0032 V	良 (Pass)
	-19.0000 V	-19.0032 V	-19.0000 V	-18.9968 V	良 (Pass)
200 V	190.000 V	189.968 V	190.000 V	190.032 V	良 (Pass)
	-190.000 V	-190.032 V	-190.000 V	-189.968 V	良 (Pass)
1000 V	1000.00 V	999.81 V	1000.00 V	1000.19 V	良 (Pass)
	-1000.00 V	-1000.19 V	-1000.00 V	-999.81 V	良 (Pass)

2. 交流電圧測定 (AC Voltage Measurement) (AC Coupling, 5 1/2 digits display, IT:100ms(SLOW1), Frequency:1kHz)

レンジ (Range)	試験値 (Tested)	規格下限 (Min)	測定値 (Data)	規格上限 (Max)	判定 (Pass/Fail)
200 mV	190.000 mV	189.710 mV	189.999 mV	190.290 mV	良 (Pass)
2000 mV	1900.00 mV	1897.10 mV	1900.02 mV	1902.90 mV	良 (Pass)
20 V	19.0000 V	18.9710 V	19.0002 V	19.0290 V	良 (Pass)
200 V	190.000 V	189.710 V	190.002 V	190.290 V	良 (Pass)
700 V	700.00 V	698.30 V	700.00 V	701.70 V	良 (Pass)

3. 直流電流測定 (DC Current Measurement) (5 1/2 digits display, IT:100ms(SLOW1))

レンジ (Range)	試験値 (Tested)	規格下限 (Min)	測定値 (Data)	規格上限 (Max)	判定 (Pass/Fail)
2000 nA	1900.00 nA	1896.75 nA	1900.00 nA	1903.25 nA	良 (Pass)
	-1900.00 nA	-1903.25 nA	-1900.02 nA	-1896.75 nA	良 (Pass)
20 µA	19.0000 µA	18.9805 µA	19.0001 µA	19.0195 µA	良 (Pass)
	-19.0000 µA	-19.0195 µA	-18.9999 µA	-18.9805 µA	良 (Pass)
200 µA	190.000 µA	189.884 µA	190.001 µA	190.116 µA	良 (Pass)
	-190.000 µA	-190.116 µA	-190.001 µA	-189.884 µA	良 (Pass)
2000 µA	1900.00 µA	1899.19 µA	1900.00 µA	1900.81 µA	良 (Pass)
	-1900.00 µA	-1900.81 µA	-1899.99 µA	-1899.19 µA	良 (Pass)
20 mA	19.0000 mA	18.9884 mA	19.0001 mA	19.0116 mA	良 (Pass)
	-19.0000 mA	-19.0116 mA	-19.0003 mA	-18.9884 mA	良 (Pass)
200 mA	190.000 mA	189.899 mA	190.002 mA	190.101 mA	良 (Pass)
	-190.000 mA	-190.101 mA	-190.002 mA	-189.899 mA	良 (Pass)
2000 mA	1900.00 mA	1897.09 mA	1899.98 mA	1902.91 mA	良 (Pass)
	-1900.00 mA	-1902.91 mA	-1899.98 mA	-1897.09 mA	良 (Pass)

4. 交流電流測定 (AC Current Measurement) (AC Coupling, 5 1/2 digits display, IT:100ms(SLOW1), Frequency:1kHz)

レンジ (Range)	試験値 (Tested)	規格下限 (Min)	測定値 (Data)	規格上限 (Max)	判定 (Pass/Fail)
200 µA	190.000 µA	189.230 µA	190.002 µA	190.770 µA	良 (Pass)
2000 µA	1900.00 µA	1892.35 µA	1900.02 µA	1907.65 µA	良 (Pass)
20 mA	19.0000 mA	18.9230 mA	19.0001 mA	19.0770 mA	良 (Pass)
200 mA	190.000 mA	189.230 mA	190.001 mA	190.770 mA	良 (Pass)
2000 mA	1900.00 mA	1892.35 mA	1900.00 mA	1907.65 mA	良 (Pass)

試験結果 Test Data

管理番号:1234567890
(Control No.)

型名 (Model No.) 7 3 5 2 A
製造番号 (Serial No.) 1 2 3 4 5 6 7 8 9

[Ach]

5. 抵抗測定 (Resistance Measurement) (2-wire, 5 1/2 digits display, IT:100ms(SLOW1))

レンジ (Range)	試験値 (Tested)	規格下限 (Min)	測定値 (Data)	規格上限 (Max)	判定 (Pass/Fail)
200	190.000	189.954	189.999	190.046	良 (Pass)
2000	1900.00	1899.71	1900.00	1900.29	良 (Pass)
20 k	19.0000 k	18.9971 k	19.0001 k	19.0029 k	良 (Pass)
200 k	190.000 k	189.957 k	190.001 k	190.043 k	良 (Pass)
2000 k	1900.00 k	1899.33 k	1900.01 k	1900.67 k	良 (Pass)
20 M	19.0000 M	18.9610 M	18.9998 M	19.0390 M	良 (Pass)
200 M	100.000 M	98.490 M	99.999 M	101.510 M	良 (Pass)

6. 周波数測定 (Frequency Measurement) (5 1/2 digits display, Gate Time:1000ms)

試験値 (Tested)	規格下限 (Min)	測定値 (Data)	規格上限 (Max)	判定 (Pass/Fail)
199.900 kHz	199.861 kHz	199.900 kHz	199.939 kHz	良 (Pass)

[Bch]

1. 直流電圧測定 (DC Voltage Measurement) (5 1/2 digits display, IT:100ms(SLOW1))

レンジ (Range)	試験値 (Tested)	規格下限 (Min)	測定値 (Data)	規格上限 (Max)	判定 (Pass/Fail)
200 mV	190.000 mV	189.972 mV	190.000 mV	190.028 mV	良 (Pass)
	-190.000 mV	-190.028 mV	-189.999 mV	-189.972 mV	良 (Pass)
2000 mV	1900.00 mV	1899.78 mV	1900.01 mV	1900.22 mV	良 (Pass)
	-1900.00 mV	-1900.22 mV	-1900.01 mV	-1899.78 mV	良 (Pass)
20 V	19.0000 V	18.9967 V	19.0000 V	19.0033 V	良 (Pass)
	-19.0000 V	-19.0033 V	-19.0000 V	-18.9967 V	良 (Pass)
200 V	190.000 V	189.967 V	190.000 V	190.033 V	良 (Pass)
	-190.000 V	-190.033 V	-190.001 V	-189.967 V	良 (Pass)

2. 直流電流測定 (DC Current Measurement) (5 1/2 digits display, IT:100ms(SLOW1))

レンジ (Range)	試験値 (Tested)	規格下限 (Min)	測定値 (Data)	規格上限 (Max)	判定 (Pass/Fail)
10 A	10.0000 A	9.9795 A	9.9999 A	10.0205 A	良 (Pass)
	-10.0000 A	-10.0205 A	-9.9999 A	-9.9795 A	良 (Pass)

3. 交流電流測定 (AC Current Measurement) (AC Coupling, 5 1/2 digits display, IT:100ms(SLOW1), Frequency:1kHz)

レンジ (Range)	試験値 (Tested)	規格下限 (Min)	測定値 (Data)	規格上限 (Max)	判定 (Pass/Fail)
10 A	10.0000 A	9.9300 A	10.0001 A	10.0700 A	良 (Pass)

4. 温度測定 (Temperature Measurement)

試験値 (Tested)	規格下限 (Min)	測定値 (Data)	規格上限 (Max)	判定 (Pass/Fail)
0.0	-3.0	0.1	3.0	良 (Pass)