

試験結果(Test Data)

型 名 (Model No.) 7 3 5 5
 製造番号 (Serial No.) 1 2 3 4 5 6 7 8 9

1、直流電圧測定 (DC Voltage Measurement) (6 1/2 digits display, IT:100ms(SLOW1))

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
200 mV	190.0000 mV	189.9712 mV	190.0001 mV	190.0288 mV	良 (Pass)
	-190.0000 mV	-190.0288 mV	-189.9998 mV	-189.9712 mV	良 (Pass)
2000 mV	1900.000 mV	1899.771 mV	1900.003 mV	1900.229 mV	良 (Pass)
	-1900.000 mV	-1900.229 mV	-1899.996 mV	-1899.771 mV	良 (Pass)
20 V	19.00000 V	18.99665 V	19.00005 V	19.00335 V	良 (Pass)
	-19.00000 V	-19.00335 V	-18.99994 V	-18.99665 V	良 (Pass)
200 V	190.0000 V	189.9685 V	190.0007 V	190.0315 V	良 (Pass)
	-190.0000 V	-190.0315 V	-189.9992 V	-189.9685 V	良 (Pass)
1000 V	1000.000 V	999.820 V	1000.009 V	1000.180 V	良 (Pass)
	-1000.000 V	-1000.180 V	-999.990 V	-999.820 V	良 (Pass)

2、交流電圧測定 (AC Voltage Measurement) (AC Coupling, 6 1/2 digits display, IT:100ms(SLOW1), Frequency:1kHz)

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
200 mV	190.0000 mV	189.7100 mV	190.0011 mV	190.2900 mV	良 (Pass)
2000 mV	1900.000 mV	1897.100 mV	1900.012 mV	1902.900 mV	良 (Pass)
20 V	19.00000 V	18.97100 V	19.00013 V	19.02900 V	良 (Pass)
200 V	190.0000 V	189.7100 V	190.0014 V	190.2900 V	良 (Pass)
700 V	700.000 V	698.300 V	700.015 V	701.700 V	良 (Pass)

3、直流電流測定 (DC Current Measurement) (6 1/2 digits display, IT:100ms(SLOW1))

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
200 mA	190.0000 mA	189.9330 mA	190.0016 mA	190.0670 mA	良 (Pass)
	-190.0000 mA	-190.0670 mA	-189.9983 mA	-189.9330 mA	良 (Pass)
2000 mA	1900.000 mA	1899.000 mA	1900.018 mA	1901.000 mA	良 (Pass)
	-1900.000 mA	-1901.000 mA	-1899.981 mA	-1899.000 mA	良 (Pass)
10 A	10.00000 A	9.98440 A	10.00020 A	10.01560 A	良 (Pass)
	-10.00000 A	-10.01560 A	-9.99979 A	-9.98440 A	良 (Pass)

4、交流電流測定 (AC Current Measurement) (AC Coupling, 6 1/2 digits display, IT:100ms(SLOW1), Frequency:1kHz)

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
200 mA	190.0000 mA	189.3300 mA	190.0022 mA	190.6700 mA	良 (Pass)
2000 mA	1900.000 mA	1892.350 mA	1900.023 mA	1907.650 mA	良 (Pass)
10 A	10.00000 A	9.94500 A	10.00024 A	10.05500 A	良 (Pass)

5、抵抗測定 (Resistance Measurement) (4-wire, 6 1/2 digits display, IT:100ms(SLOW1))

レンジ(Range)	試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
200 Ω	190.0000 Ω	189.9540 Ω	190.0025 Ω	190.0460 Ω	良 (Pass)
2000 Ω	1900.000 Ω	1899.704 Ω	1900.026 Ω	1900.296 Ω	良 (Pass)
20 kΩ	19.00000 kΩ	18.99704 kΩ	19.00027 kΩ	19.00296 kΩ	良 (Pass)
200 kΩ	190.0000 kΩ	189.9590 kΩ	190.0028 kΩ	190.0410 kΩ	良 (Pass)
2000 kΩ	1900.000 kΩ	1899.370 kΩ	1900.029 kΩ	1900.630 kΩ	良 (Pass)
20 MΩ	19.00000 MΩ	18.96140 MΩ	19.00030 MΩ	19.03860 MΩ	良 (Pass)
200 MΩ	180.0000 MΩ	177.2940 MΩ	180.0031 MΩ	182.7060 MΩ	良 (Pass)

6、周波数測定 (Frequency Measurement) (6 1/2 digits display, Gate Time:1000ms)

試験値(Tested)	規格下限(Min)	測定値(Data)	規格上限(Max)	判定(Pass/Fail)
199.9000 kHz	199.8601 kHz	199.9032 kHz	199.9399 kHz	良 (Pass)